

04/12/00
1c520 U.S. PTO

09/547945 - 04/12/00

4-13-00

A

Please type a plus sign (+) inside this box → ☐

PTO/SB/05 (4/98)
Approved for use through 09/30/2000. OMB 0651-0032
Patent and Trademark Office U.S. DEPARTMENT OF COMMERCE
Under the Paperwork Reduction Act of 1995, no persons are required to respond to a collection of information unless it displays a valid OMB control number.

UTILITY PATENT APPLICATION TRANSMITTAL

(Only for new nonprovisional applications under 37 C.F.R. § 1.53(b))

Attorney Docket No. 1930
First Inventor or Application Identifier Beatty et al.
Title Electrophysiology Mapping System
Express Mail Label No. EK58464313US

APPLICATION ELEMENTS

See MPEP chapter 600 concerning utility patent application contents

ADDRESS TO: Assistant Commissioner for Patents
Box Patent Application
Washington, DC 20231

1. ☒ * Fee Transmittal Form (e.g., PTO/SB/17)
(Submit an original and a duplicate for fee processing)
2. ☒ Specification [Total Pages 40]
(preferred arrangement set forth below)
 - Descriptive title of the Invention
 - Cross References to Related Applications
 - Statement Regarding Fed sponsored R & D
 - Reference to Microfiche Appendix
 - Background of the Invention
 - Brief Summary of the Invention
 - Brief Description of the Drawings (if filed)
 - Detailed Description
 - Claim(s)
 - Abstract of the Disclosure
3. ☒ Drawing(s) (35 U.S.C. 113) [Total Sheets 17]
4. Oath or Declaration [Total Pages]
 - a. ☐ Newly executed (original or copy)
 - b. ☐ Copy from a prior application (37 C.F.R. § 1.63(d))
(for continuation/divisional with Box 16 completed)
 - i. ☐ DELETION OF INVENTOR(S)
Signed statement attached deleting
inventor(s) named in the prior application,
see 37 C.F.R. §§ 1.63(d)(2) and 1.33(b).

* NOTE FOR ITEMS 1 & 13: IN ORDER TO BE ENTITLED TO PAY SMALL ENTITY
FEES, A SMALL ENTITY STATEMENT IS REQUIRED (37 C.F.R. § 1.27), EXCEPT
IF ONE FILED IN A PRIOR APPLICATION IS RELIED UPON (37 C.F.R. § 1.28)

5. ☐ Microfiche Computer Program (Appendix)
6. Nucleotide and/or Amino Acid Sequence Submission
(if applicable, all necessary)
 - a. ☐ Computer Readable Copy
 - b. ☐ Paper Copy (identical to computer copy)
 - c. ☐ Statement verifying identity of above copies

ACCOMPANYING APPLICATION PARTS

7. ☐ Assignment Papers (cover sheet & document(s))
8. ☐ 37 C.F.R. § 3.73(b) Statement ☐ Power of
(when there is an assignee) ☐ Attorney
9. ☐ English Translation Document (if applicable)
10. ☐ Information Disclosure ☐ Copies of IDS
Statement (IDS)/PTO-1449 ☐ Citations
11. ☐ Preliminary Amendment
12. ☒ Return Receipt Postcard (MPEP 503)
(Should be specifically itemized)
13. ☐ * Small Entity ☒ Statement filed in prior application,
Statement(s) ☐ Status still proper and desired
(PTO/SB/09-12)
14. ☐ Certified Copy of Priority Document(s)
(if foreign priority is claimed)
15. ☒ Other: Check for \$345
Unsigned Declaration

16. If a CONTINUING APPLICATION, check appropriate box, and supply the requisite information below and in a preliminary amendment:

☐ Continuation ☒ Divisional ☐ Continuation-in-part (CIP) of prior application No. 09 / 005,105
Prior application information Examiner Winakur, E Group / Art Unit 3736

For CONTINUATION or DIVISIONAL APPS only: The entire disclosure of the prior application, from which an oath or declaration is supplied
under Box 4b, is considered a part of the disclosure of the accompanying continuation or divisional application and is hereby incorporated by
reference. The incorporation can only be relied upon when a portion has been inadvertently omitted from the submitted application parts.

17. CORRESPONDENCE ADDRESS

☐ Customer Number or Bar Code Label (Insert Customer No. or Attach bar code label here) or ☒ Correspondence address below

Name	Robert C. Beck			
	Beck & Tysver			
Address	1011 First Street South Suite 440			
City	Hopkins	State	MN	Zip Code 55343
Country	USA	Telephone	952-933-3412	Fax 952-933-3049

Name (Print/Type)	Robert C. Beck	Registration No. (Attorney/Agent)	28,184
Signature	<i>Robert C. Beck</i>	Date	4/12/00

Burden Hour Statement This form is estimated to take 0.2 hours to complete. Time will vary depending upon the needs of the individual case. Any
comments on the amount of time you are required to complete this form should be sent to the Chief Information Officer, Patent and Trademark Office,
Washington, DC 20231. DO NOT SEND FEES OR COMPLETED FORMS TO THIS ADDRESS. SEND TO Assistant Commissioner for Patents,
Box Patent Application, Washington, DC 20231

ELECTROPHYSIOLOGY THERAPY CATHETER1. Cross Referenced to Related Cases

This is a Divisional Application of U.S. Serial No. 09/005,105, filed January 9, 1998, entitled "Electrophysiology Mapping System;" which was a Continuation-In-Part of U.S. Serial No. 08/387,832, officially filed May 26, 1995 and entitled "Endocardial Mapping System and Catheter Probe"; which is a national stage application based upon international application PCT/US93/09015, filed September 23, 1993; which in turn is a Continuation-In-Part of both U.S. Serial No. 07/950,448, filed September 23, 1992 (now U.S. Patent 5,291,549) and U.S. Serial No. 07/949,690, also filed September 23, 1992 (now U.S. Patent 5,311,866). The parent application, Serial No. 08/387,832, is incorporated by reference herein.

2. Field of the Invention

The parent invention relates to electrophysiology apparatus which is used to measure and to visualize electrical activity occurring in a patient's heart. The system can display both a visual map of the underlying electrical activity originating in a chamber of a patient's heart and the location of a therapy catheter located within a heart chamber. The electrophysiology apparatus includes several subsystems including: a therapy catheter system, a measurement catheter system and a computer based signal acquisition, control and display system.

3. Background of the Invention

Many cardiac tachyarrhythmias are caused by conduction defects which interfere with the normal propagation of electrical signals in a patient's heart.

- 5 These arrhythmias may be treated electrically, pharmacologically or surgically. The optimal therapeutic approach to treat a particular tachyarrhythmia depends upon the nature and location of the underlying conduction defect. For this reason electrophysiologic mapping is used
- 10 to explore the electrical activity of the heart during a tachyarrhythmic episode. The typical electrophysiologic mapping procedure involves positioning an electrode system within the heart. Electrical measurements are made which reveal the electrical propagation of activity in the heart.
- 15 If ablation is the indicated therapy then a therapy catheter is positioned at the desired location within the heart and energy is delivered to the therapy catheter to ablate the tissue.

- There are numerous problems associated with these
- 20 electrophysiologic diagnostic and therapeutic procedures. First the testing goes on within a beating heart. The motion of the diagnostic catheter and treatment catheter can injure the heart and provoke bouts of arrhythmia which interfere with the collection of diagnostic information.
- 25 During the delivery of ablation therapy it is common to use fluoroscopic equipment to visualize the location of the

catheters. Many physicians are concerned about routine occupational exposure to X-rays. In addition, the traditional mapping techniques do not provide a high resolution view of the electrical activity of the heart which makes it difficult to precisely locate the source of the arrhythmia.

Summary

The electrophysiology apparatus of the invention is partitioned into several interconnected subsystems. The measurement catheter system introduces a modulated electric field into the heart chamber. The blood volume and the moving heart wall surface modify the applied electric field. Electrode sites within the heart chamber passively monitor the modifications to the field and a dynamic representation of the location of the interior wall of the heart is developed for display to the physician. Electrophysiologic signals generated by the heart itself are also measured at electrode sites within the heart and these signals are low pass filtered and displayed along with the dynamic wall representation. This composite dynamic electrophysiologic map may be displayed and used to diagnose the underlying arrhythmia.

A therapy catheter system can also be introduced into the heart chamber. A modulated electrical field delivered to an electrode on this therapy catheter can be

within the same chamber. The potential sensed from this current may be processed to determine the relative or absolute location of the therapy catheter within the chamber.

5 These various processes can occur sequentially or simultaneously several hundred times a second to give a continuous image of heart activity and the location of the therapy device.

10 Brief Description of the Drawings

 An exemplary and illustrative form of the invention is shown in the drawings and identical reference numerals refer to equivalent structure throughout.

 FIG. 1 is a schematic block diagram of the
15 electrophysiology apparatus;

 FIG. 2 is a block diagram representing the partitioning of the electrophysiology apparatus;

 FIG. 3 is a diagram of an illustrative balloon electrode set implementation of the measurement catheter
20 and a therapy catheter;

 FIG. 4 is a schematic diagram of an illustrative basket electrode set implementation of the measurement catheter;

 FIG. 5 is a flow chart showing the wall surface
25 generation process;

 FIG. 6 is a schematic diagram of a row of

electrodes of the balloon catheter and their use in measuring distance to the heart chamber wall;

FIG. 7 is a screen display representing the motion of the cardiac wall surface;

5 FIG. 8 is a schematic block diagram of the portion of the electrophysiology apparatus which implements the body orientation generation process;

FIG. 9 is a flow charting showing the body orientation generation process;

10 FIG. 10 is a flow chart showing the wall electrogram generation process;

FIG. 11 is a representative screen display showing wall electrogram information;

15 FIG. 12 is a representative screen display showing wall electrogram information;

FIG. 13 is a representative screen display showing wall electrogram information;

FIG. 14 is a flow chart showing the site electrogram generation process; and

20 FIG. 15 is a flow chart showing the movable electrode location process.

FIG. 16 is a schematic block diagram of the therapy catheter system;

25 FIG. 17 is a schematic diagram of the laser delivery embodiment of the therapy catheter;

FIG. 18 is a schematic diagram of a microwave

delivery embodiment of the therapy catheter;

FIG. 19 is a schematic diagram of a chemical delivery embodiment of the therapy catheter; and

FIG. 20 is a schematic diagram of the angioplasty catheter embodiment of the therapy catheter.

Detailed Description

FIG. 1 shows the electrophysiologic apparatus 10 connected to a patient 12. In a typical procedure a monitoring catheter system 14 is placed in the heart 16 to generate a display of the electrical activity of the heart 16. After diagnosis a therapy catheter 18 may be inserted into the heart to perform ablation or other corrective treatment.

The monitoring catheter 14 has a proximal end 20 which may be manipulated by the attending physician, and a distal end 22 which carries a monitoring catheter electrode set 44. In general the distal end 22 of the monitoring catheter 14 will be relatively small and will float freely in the heart chamber. The therapy catheter 18 has a distal end 24 which carries a therapy catheter electrode set 46. The therapy catheter also has proximal end 26 which can be manipulated by the attending physician.

The electrode sets located on the catheters are coupled to an interface system 28, through appropriate cables. The cable 30 connects the monitoring catheter

electrode set 44 to the interface system 28 while cable 32 connects the therapy catheter electrode set 46 to the interface system 28. The interface system 28 contains a number of subsystems which are controlled by a computer 34.

5 The data collected by the interface system 28 is manipulated by the computer 34 and displayed on a display device 36. Surface electrodes represented by electrode 40 may also be coupled to the electrophysiology apparatus 10 for several purposes via an appropriate cable 42. A

10 therapy generator 38 is connected to the therapy catheter electrode 60 and to the therapy surface ground 70, through the interface system 28. The skin surface electrode cable 42 couples the ECG surface electrodes 74 to the ECG system 39, which may be a subsystem of interface system 28.

15 FIG. 2 is a schematic diagram showing an illustrative segmentation of the electrode sets and their electrical connections to subsystems in the electrophysiology apparatus 10. For example the monitoring electrode set 44 contains a subset of passive electrodes 48
20 which are connected to a signal conditioner 50. The monitoring electrode set 44 also contains a subset of active electrodes 52 which are connected to a signal generator 54 through a switch 59. The signal generator 54 is controlled by the computer 34. In operation, the signal
25 generator 54 generates a burst of (4800 Hz for example) signals which are supplied to the active electrode set 52.

003740" 0462650
This energy sets up an electric field within the heart 16
chamber. The electrical potentials present on the passive
electrode set 48 represent the summation of the underlying
electrophysiological signals generated by the heart and the
5 field induced by the burst. The signal conditioner 50
separates these two components. The preferred technique is
to separate the signals based upon their frequency.

The high pass section 56 of the signal
conditioner extracts the induced field signals as modulated
10 by the blood volume and the changing position of the
chamber walls 125. First, the signals are amplified with a
gain of approximately 500 from passive electrodes 48 with
amplifier 151. Next, the signals are high pass filtered at
roughly 1200 Hz by filter 153. Then the 4800 Hz signal is
15 extracted by demodulator 155. Finally, the individual
signals are converted to digital format by the analog to
digital converter 157 before being sent to the computer 34.

The low pass section 58 of the signal conditioner
50 extracts physiologic signals. First, signal drift is
20 reduced with a 0.01 Hz high pass filter 143. Next, a
programmable gain amplifier 145 amplifies the signals.
Then a low pass filter 147 removes extraneous high
frequency noise and the signal from the induced field.
Finally, the physiologic signals are converted to digital
25 format by the analog to digital converter 149 before being
sent to the computer 34.

The therapy catheter electrode set 46 includes at least one therapy delivery electrode 60, and preferably one or more monitoring electrodes 62, and one or more locator electrodes 68. The therapy delivery electrode 60 cooperates with the ground electrode 70, which is generally a skin patch electrode, to deliver ablation energy to the heart. These electrodes are coupled to the ablation energy generator 38 which is shown as an RF current source. A locator electrode 68 is provided which is preferably proximate the delivery electrode 60, but can be a separate electrode site located near the distal end 24 of the therapy catheter 18. This electrode site is coupled with an active electrode 52 through a switch 59 to the signal generator 54. In use, the electric field coupled to the therapy catheter 18 permits the physician to track and visualize the location of the locator electrode 68 on the display device 36. The therapy catheter electrode set 46 can also be used to monitor the physiologic signals generated at the chamber wall 125 by a low pass signal conditioner 141 which is similar to the low pass section 58 of the signal conditioner 50. These digitized signals are then sent to the computer 34.

At least one electrode pair 119 of surface electrodes 40 are also coupled to the signal generator 54 through switch 59. Each electrode 89 and 115 are placed opposite each other on the body surface with the heart 16

in-between them. The induced field is sensed by passive electrodes 48 and conditioned by the high pass section 56 of the signal conditioner 50. This field helps the computer 34 align or orient the passive electrodes 48 to the body for better visualization of the heart on the monitor 36.

The ECG subsystem 39 accepts signals from standard ECG skin electrodes 74. It also contains a low pass section similar to the low pass section 58 of signal conditioner 50. In general, the passive electrode set 48 and active electrode set 52 will reside on a single catheter, however it should be recognized that other locations and geometries are suitable as well. Both basket and balloon devices are particularly well suited to this application.

FIG. 3 shows an electrode configuration on a balloon catheter 94 which has an inflatable balloon 96 which underlies an array or set of passive electrodes 48 typified by passive electrode 72. These passive electrodes 48 can be organized into rows, typified by row 123, and columns, typified by column 121. A pair of active excitation electrodes 52 are typified by proximal electrode 92 and distal electrode 98. The balloon catheter 94 configuration can be quite small in comparison with the basket catheter 80 configuration. This small size is desirable both for insertion into and for use in a beating

heart 16.

FIG. 3 also shows a movable, reference or therapy catheter system 18. This catheter is shown lying along the interior surface 125 of the heart 16. A pair of electrodes shown as delivery electrode 60 and reference electrode 62 are located a fixed distance apart on the catheter body 64. This auxiliary catheter may be used to supply ablation energy to the tissue during therapy. This therapy catheter 18 may be used with either the basket catheter 80 configuration or the balloon catheter 94 configuration.

FIG. 4 shows an electrode configuration on a basket catheter 80. The limbs of the basket 80, typified by limb 82 carry multiple passive electrode sites typified by electrode 84. A pair of active excitation electrodes are shown on the central shaft 86 of the basket 80 as indicated by excitation electrode 88. The basket catheter 80 electrodes lie gently against the interior surface 125 of the heart 16 urged into position by the resilience of the limbs. The basket catheter 80 permits unimpeded flow of blood through the heart during the mapping procedure which is very desirable. This form of catheter also places the electrodes into contact with the heart chamber wall 125 for in-contact mapping of the physiologic potentials of the heart 16.

Returning to FIG. 1 and FIG. 2 these figures show one illustrative partitioning of system functions. In use,

the movable electrode location process (MELP).

WALL SURFACE GENERATION PROCESS

FIG. 5 is a flow chart describing the method used to generate the "wall surface" of the interior of the heart 16. The step-wise processes are presented with certain physical parameters which are either known in advance by computation or are measured. This knowledge or information is shown in block 53, block 55 and block 57. The WSGP process begins at block 41 with the insertion of the monitoring catheter 14 in the heart 16. This catheter 14 places an array of electrodes 44 in a heart 16 chamber. This array must have both passive measurement electrode sites 48 and active interrogation electrode sites 52 located in a known position. The process enters a measurement and display loop at block 43 where an interrogation pulse burst is generated by the signal generator 54 seen in FIG. 2. These pulses are generated first with the current source at site 92 and the current sink at site 98 and second with the current source at site 98 and the sink at site 92 as seen in FIG. 3. At block 45 the signal conditioner 50 uses information on the frequency and timing of the interrogation current from block 53 to demodulate the signals and analog to digital convert the signals received at the passive measurement electrodes 48. At block 47 the information from block 55 is used. This

information includes both the current strength of the interrogation pulse and the location of the interrogation source and sink electrodes. Impedance is voltage divided by current. The voltage offset caused by the location of the current source can be reduced by the two measurements of opposite polarity. This information is used to determine the impedance which the chamber and the blood contained in that chamber imposes on the field generated by the interrogation current. The knowledge from block 57 is used next. Block 49 determines how the heart chamber tissue, which has roughly three times the impedance of blood, in combination with the type of electrode array affects the field generated by the interrogation electrodes.

In a system as shown as the basket in FIG. 4 the blood effects the impedance directly as the field is propagated from the interrogation electrodes to the measurement electrodes. In general, if a point current course is used within a chamber the inverse of the measured voltage is proportional to the square root of the distance from the source. With the distance from each electrode 84 to both excitation electrodes 88 computed from the measured voltage and the known location of the electrodes 84 relative to each other, the locations of each electrode 84 can be determined.

In a system as shown in FIG. 3 the impedance of

the field generated within the blood volume is modulated by the position of the walls 125, with their higher impedance, with respect to the location relative to the measurement electrodes. Using this knowledge and the measurements from
5 block 47 the distance from the interrogation electrodes to the heart chamber wall 125 is determined at a point normal to the field generated by the active interrogation electrodes 52.

The passive electrodes 48 on the balloon catheter
10 94 can be positioned in rows 123 and columns 121 with the columns in a line from the top of the balloon 96 near active electrode 92 to the bottom of the balloon 96 near active electrode 98. In a preferred embodiment three configurations are possible: 8 rows and 8 columns, 7 rows
15 and 9 columns, and 6 rows and 10 columns. In each such embodiment the measurements from any row 123 are treated independently. Using the 8 row, 8 column embodiment as an example, 8 measurements of distance are taken for any selected row of electrodes, giving a total of
20 64 measurements.

FIG. 6 is a schematic drawing of the embodiment required to measure the distance 129 from the centroid 127 of the balloon 96 through the passive electrode 131 to the heart chamber wall 125. The passive electrode 131 is one
25 of eight electrodes on a row of electrodes 123. Starting with electrode 131 and labeling it as electrode A, the

other electrodes on the row 123 are labeled B, C, D, E, F, G and H by proceeding around the balloon 96 in a clockwise direction. The measurements of impedance "I" at these electrodes are thus labeled I_A , I_B , I_C , I_D , I_E , I_F , I_G and I_H .

5 To compute the distance 129 in the direction of electrode 131 the following equation is computed:

$$\ln(D_A) = c_0 + c_1 \ln(I_A) + c_2 \ln(I_B) + c_3 \ln(I_C) + c_4 \ln(I_D) + c_5 \ln(I_E) + c_4 \ln(I_F) + c_3 \ln(I_G) + c_2 \ln(I_H)$$

where D_A is the desired distance 129 and c_0 through c_5 are
10 optimized parameters. A typical vector of these parameters is $(c_0, c_1, c_2, c_3, c_4, c_5) = (3.26, -.152, -.124, -.087, -.078, -.066)$.

Once the distance 129 in the direction of electrode 131 is determined then the computation can be
15 redone by shifting this direction clockwise one electrode, relabeling electrodes A through H and solving the above equation again. Once the distances for this row of electrodes 123 are determined then the next row distances are determined in the same way until the distances at all
20 64 electrodes are determined.

Returning to FIG. 5, with multiple wall locations in space determined by this method, a model of the chamber wall 125 shape can be created in block 51. Various techniques for creating a shape are possible, including
25 cubic spline fits, and best fit of an ellipsoid. The positions of the active electrodes 52 and the passive

electrodes 48 relative to the heart 16 chamber walls are also determined at this point. The loop continues as the method moves back to block 43. This loop continues at a rate fast enough to visualize the real-time wall motion of the heart chamber, at least at twenty times per second.

There are numerous display formats or images which can be used to present the dynamic endocardial wall surface to the physician. It appears that one of the most useful is to unfold the endocardial surface and project it onto a plane. Wire grid shapes representing a perspective view of the interior of the heart chamber are useful as well. It appears that each individual physician will develop preferences with respect to preferred output image formats. In general, different views of the endocardial surface will be available or may be used for diagnosis of arrhythmia and the delivery of therapy. One distinct advantage of the present invention is that the image of the heart wall is not static or artificial. In this system the image is a measured property of the heart wall, and is displayed in motion.

FIG. 7 shows two separate frames of the dynamic representation of the heart wall. Wire frame 71 shows the heart at systole while wire frame 73 shows the heart at diastole. Path arrow 75 and path arrow 77 represent the dynamic cycling through several intermediate shapes between the systole and diastole representation. These views are

useful as they indicate the mechanical pumping motion of the heart to the physician.

BODY ORIENTATION GENERATION PROCESS

5 FIG. 8 is a schematic drawing of the apparatus required to perform the body orientation generation process. It shows a patient 12 with at least one pair 119 of skin electrodes 40 attached to the body surface in a stationary position on the body and in a known
10 configuration. These electrodes are typified by example surface electrodes 89 and 115 each of which could be an ECG electrode 74, an RF generation current sink electrode 70, or another electrode specifically dedicated to the BOGP. Ideally, electrode 89 and 115 are opposite one another on
15 the body with the heart 16 directly in between them. This pair of electrodes is attached to the signal generator 54 through the switch 59 via an appropriate cable 117. The distal end 22 of monitoring catheter 14 is situated in the heart 16 where the passive electrodes 48 can measure the
20 signals generated across the electrode 89 and electrode 115.

FIG. 9 is a flow chart describing the method used to align the wall surface representation of the WSGP to the body orientation. The process begins at step 101 where the
25 monitoring catheter 14 with a set of passive electrodes 48 is inserted into heart 16 chamber and a pair of surface

electrodes 119 are attached at a known position on the body
12. The process begins cycling at step 102 where the
signal generator 54 generates a signal across the skin
electrode 89 and skin electrode 115. At step 103 the
5 voltage created by the signal generator 54 is measured from
passive electrode 48 by the high pass section 56 of the
signal conditioner 50 by using the information from block
110 which includes the frequency and timing of the field
generated by the signal generator 54. This voltage
10 information is stored in an array "Y".

At step 104 a regression analysis is performed
which creates a vector which lines up with the field
generated in step 103. This regression method is the same
whether a basket catheter as shown in FIG. 4 or a balloon
15 catheter as shown in FIG. 3 is used. The location of each
passive electrode 48 is provided to the method by block
110. This information comes from different sources in each
case however. In the case of a basket catheter 80 these
three dimensional electrode locations come from the WSGP.
20 In the case of the balloon catheter 94 these three
dimensional electrode locations are known a priori. In
each case they are saved in an array "X". The regression
to compute the orientation vector uses the standard
regression equation for the computation of a slope:

25
$$b = \Sigma xy / \Sigma x^2$$

where "X" is the array of electrode locations, "Y" is the

array of measured voltages and "b" is the orientation vector. If more than one pair of skin electrodes are used then an orthogonal set of orientation vectors can be created and any rotation of the monitoring catheter 14
5 relative to the body 12 can be detected.

In step 105 the information on the location of the chamber walls 125 from the WSGP 109 can be used to create a three dimensional model of the heart 16 chamber as seen in FIG. 7. By combining this model with the computed
10 orientation from step 104 and the known location of the skin electrodes 108 this representation can be shown in a known orientation relative to the body in step 106. In step 107 a specific orientation such as typical radiological orientations RAO (right anterior oblique), LAO
15 (left anterior oblique), or AP (anterior/posterior) can be presented. By repeatedly showing this view a dynamic representation can be presented which matches the view shown on a standard fluoroscopic display. Thus such an image can be presented without the need for using ionizing
20 radiation.

WALL ELECTROGRAM GENERATION PROCESS

FIG. 10 is a flow chart describing the wall electrogram generation process (WEGP). This process begins
25 at block 61 when a monitoring catheter 14 with an array of passive measurement electrodes 48 is placed in a heart

chamber 16 and deployed. The process enters a loop at block 63. The frequency of the interrogation pulses generated by the signal generator 54 is provided by block 85. With this knowledge the low pass filter section 58 of the signal conditioner 50 measures the voltage at frequencies lower than the generated interrogation pulses. Typically the highest frequency of the biopotentials is 100 Hz but can be as high as 250 Hz.

In the case of a basket system as seen in FIG. 4 the measurements are contact voltages from the chamber wall 125 tissue contacting the electrodes 84.

In the case of a balloon system as seen in FIG. 3 the measurements are measurements of the field generated throughout the blood volume by the tissue on the chamber wall 125. At step 65, a model of the array boundary and the chamber wall 125 boundary is created from the information in block 87. This information includes the location of the passive electrodes 48 on the array and the chamber wall 125 locations from the WSGP.

In the case of a basket system as seen in FIG. 4, the array boundary and the chamber wall 125 boundary are the same since they are in contact. The locations are determined in three-dimensional space of the sites on the chamber wall where potentials are measured.

In the case of the balloon system as seen in FIG. 3, the array boundary and the chamber wall 125

boundary are different. During step 65, locations are generated in three-dimensional space of the sites on the chamber wall where potentials are to be determined.

At step 66, the potentials are projected on to the sites on the chamber wall specified in step 65. In the case of a basket system as seen in FIG. 4, the measured potentials are assigned to these sites.

In case of a balloon system as seen in FIG. 3, a three dimensional technique such as those typically used in field theory is used to generate a representation of the three dimensional field gradients in the blood volume of the heart chamber. Two examples of appropriate techniques are a spherical harmonics solution to Laplace's equation, and boundary element analysis. A more detailed description of spherical harmonics is given in the parent disclosure which is incorporated by reference herein.

For the boundary element method in the mapping system of the invention, the voltage is measured at the passive electrodes 48 on the probe or balloon catheter 94. From the voltage at the electrodes on the probe and the knowledge that the probe is nonconducting, the voltage and normal current at a previously selected set of nodes on the endocardial surface 125 are determined by the boundary element method in the following manner.

It is known that the voltage in the blood pool between the probe and the endocardium satisfies Laplace's

equation that states that the net current flow across any specific boundary is zero. To find the voltage and/or normal current on the endocardium, one must find the solution of Laplace's equation in the blood pool and
5 calculate the values of this solution on the endocardium. Standard finite element and finite difference methods can be used to find the solution to Laplace's equation, but they have large computational overhead for generating and keeping track of a three-dimensional grid in the whole
10 blood pool. In the mapping system of the invention, Laplace's equation is solved by the boundary element method, a specialized finite element method that permits one to restrict the calculations to the two-dimensional probe and endocardial surfaces (and not have to deal with
15 calculations over the blood pool between these two surfaces). In order to create an accurate map of the endocardial voltage and/or normal current based on the voltage information from a limited number of electrodes on the probe, the system uses a higher-order version of the
20 boundary element method. This system currently uses bicubic splines to represent the probe and endocardial surfaces and bilinear elements and bicubic splines to represent the voltage and the normal current on these surfaces.
25 The boundary element method consists of creating and solving a set of linear equations for the voltage and

the normal current on the endocardium based on the voltage measurements at the electrodes on the probe. Each of the elements in the matrices that are involved in this set consists of two-dimensional integrals, which are calculated
5 by numerical and analytical integration.

Using Laplace's equation with data given on the probe is a so-called "ill-posed" problem. For such problems, all solution procedures, including the boundary element method, are ill conditioned, that is, small errors
10 in the measured voltage on the probe surface can result in large errors in the calculated voltage and/or normal current on the endocardium. To minimize the errors on the endocardium, options for regularization or constraints have been included in the software code. For example: the user
15 can choose parameters that cause the code to add equations for known or expected values of the voltage and/or normal current on the endocardium. This capability is often but not exclusively used to add equations that take into account the voltage and/or normal current of the map of the
20 previous instant(s) in time (the previous "frame(s)"). This process uses historical data from the previous frame to constrain the values subsequently computed.

The solution of the set of the boundary element equations and regularizing equations (if any) is normally
25 accomplished by singular value decomposition but there is an option to solve the linear system by decomposition

(Gaussian elimination) or direct or inherent methods. When singular value decomposition is used, there is an option to turn off the influence of high-frequency errors (that is, do a type of regularization) by setting various small
5 singular values to zero, the result of which can be an increase in the accuracy of the calculated voltage and normal current on the endocardium.

In block 67, a large number of points are calculated on the three-dimensional chamber surface 125.
10 In the case of a basket catheter as seen in FIG. 4, this is done through interpolation using bilinear or bicubic splines. In the case of a balloon catheter as seen in FIG. 3, this can be done either by using the model, such as the boundary element method or spherical harmonics to generate
15 more points. Alternatively, bilinear or bicubic splines can be used to interpolate between a smaller number of points.

In block 69 a representation of the electrical potentials on the surface 125 are used to display the
20 patterns. These types of displays include color maps, maps of iso-potential lines, maps of potential gradient lines and others. The electro-physiologic information is reconstructed on the dynamic wall surface 125. In general the measured electrical activity is positioned by the WSGP
25 at the exact location which gives rise to the activity. The high resolution of the system creates an enormous

amount of information to display. Several techniques may be used to display this information to the physician. For example the electrogram data can be shown in false color gray-scale on a two dimensional wall surface

5 representation. In this instance areas of equal potential areas are shown in the same color. Also a vectorized display of data can be shown on a wire grid as shown in FIG. 11 where the distance between any two dots typified by dot pair 91 and 93 represent a fixed potential difference.

10 The more active electrical areas show clusters of dots. In a dynamic display the dot movement highlights areas of greater electrical activity. In FIG. 12 gradient lines typified by line 135 represent the change in potential over the chamber wall surface. Those areas with the largest
15 change per unit area have the longest gradient lines oriented in the direction of steepest change. In FIG. 13 iso-potential lines typified by line 95 represent equal electrical potential. In this representation the closeness of lines represents more active electrical areas.

20

SITE ELECTROGRAM GENERATION PROCESS

FIG. 14 is a flow chart of the site electrogram generation process (SEGP). This process is used to extract and display a time series representation of the electrical
25 activity at a physician selected site. FIG. 13 shows a site 97 that has been selected and a time series

MOVABLE ELECTRODE LOCATION PROCESS

FIG. 15 is a flow-chart of the movable electrode location process (MELP). It begins at block 11 when a catheter with an array of passive measurement electrodes 48 and active electrodes 52 is placed in a heart 16 chamber and deployed. At block 13 a second catheter 18 with at least one electrode is introduced into the same chamber. The process enters a loop at block 15 where the signal generator 54 generates a carrier current between the movable location electrode 68 and an active electrode 52. At block 17 the high pass section 56 of signal conditioner 50, using the frequency and timing information of the location signal from block 29, produces measured voltages from the passive measurement electrodes 48. At block 19 the information from block 27 is used to determine the location of the electrode 68 where the location current is generated. This information includes the strength of the generated location current, the impedances of blood and tissue, the location of the active electrode 52 in use and the location of all the passive measurement electrodes 48. One method for using this information would entail performing a three dimensional triangulation of the point source location signal using four orthogonal passive electrode 48 sites. The implementation of step 19 is the same both for the case of a basket system as seen in FIG. 3 and for the case of a balloon system as seen in FIG. 4. In

this preferred implementation, two data sets are acquired closely spaced in time such that they are effectively instantaneous relative to the speed of cardiac mechanical activity. Alternatively, the data sets could be acquired
 5 simultaneously, by driving signals at two different frequencies, and separating them electronically by well known filtering means.

The first data set is acquired by driving the current carrier from the location electrode 68 to a first
 10 sink or active electrode as typified by electrode 98. This electrode is at a known location on the body of the monitoring catheter 14 relative to the array of passive electrodes 48. The location of this first sink electrode is ideally displaced distally from the centroid 127 of the
 15 array of electrodes by at least 25 millimeters. A second data set is then acquired by driving the current from the location electrode 68 to a second active electrode 92, located ideally at least 25 millimeters proximally from the centroid 127 of the array of electrodes.

20 The location algorithm is performed by minimizing the following equation:

$$\sum_{i=1}^n \left(\frac{k}{\left(\vec{R}_i - \vec{R}_L \right)^{0.5}} - V_{pi_1} - b_1 - \frac{k}{\left(\vec{R}_i - \vec{R}_{S_1} \right)^{0.5}} \right)^2 + \left(\frac{k}{\left(\vec{R}_i - \vec{R}_L \right)^{0.5}} - V_{pi_2} - b_2 - \frac{k}{\left(\vec{R}_i - \vec{R}_{S_2} \right)^{0.5}} \right)^2$$

Where n is the number of array electrodes, where k , b_1

and b_2 are fitting parameters, V_{pi} are the potentials measured from each i^{th} electrode 72, R_i is a vector from the origin (centroid of the array of electrodes 96) to the i^{th} probe electrode 72, R_L is the "location vector",
5 or three dimensional location to be solved for in the minimization, and R_{s1} , R_{s2} are the location vectors of the active sink electrodes (eg. 92 and 98) which are known at locations on the axis of the array of passive electrodes 48.

10 Additional data sets could be incorporated, following the same logic as above. Each additional squared parenthetical term requires the probe data set V_{pi} , another 'b' fitting term, and the particular active sink electrode 52 vector R_s used during the acquisition
15 of that data set. If the sink electrode 52 is far enough away, for example using a right leg patch electrode, the fourth term in the squared expression for that data set may be deleted as R_s becomes very large.

It is also noted that the method does not
20 **require** two data sets. The first squared expression in the above expression (requiring only data set V_{pi1}) may be sufficiently accurate.

The non-linear least squares minimization may be performed on the above summation by any of several
25 well-known methods. The Levenberg-Marquardt method has

been used in practice to accomplish this with efficient and robust results. Nominal values for k and b are 70 and 0 respectively, when normalizing the potential values obtained as if the current source were 1 ampere.

5 The number of parameters in the minimization for the above expression are six: k , b_1 , b_2 , and the x , y , and z coordinates of vector R_L (assuming a cartesian coordinate system with origin at the center of the array of electrodes 96).

10 At step 21 a model of the heart 16 chamber wall is generated from the information provided from the WSGP 25. Such a model can be represented on a display 36 in a manner typified in FIG. 6. Once this surface is rendered, within this surface a second figure
15 representing the distal end of the monitoring catheter 14 can be presented. In this way, the full three dimensional geometry of the chamber and the array catheter can be presented.

In step 23 this geometry is updated repeatedly
20 to provide a dynamic view of the chamber, the monitoring catheter 18, along with a representation of the distal end 24 of the therapy catheter 18. If this is then combined with the electrical potentials generated by the WEGP, the therapy catheter can be moved to an electrical
25 site of interest represented by a point in three

dimensional space.

CALIBRATION PROCESS

Calibration of the system to insure that
5 physical dimensions are accurately scaled is not a
necessity for use of the system in a diagnostic or
therapeutic setting. However, the availability of heart
geometry in real time can permit various hemodynamic
measurements to be made and displayed to the physician
10 as well. These measurements include systolic time
intervals, stroke volume and cardiac output.
Calibration, where desired, requires at least two
electrodes 60 and 62 a known distance apart placed along
the inner-surface of the heart chamber 16, as shown in
15 FIG. 3. In general the two electrode sites will each be
coupled to the location signal generator 54. The MELP
of FIG. 15 can be calibrated by scaling the calculations
50 the distance between computed locations match the
known distance apart of the two electrodes 60 and 62.
20 Since the electrodes 60 and 62 are positioned on the
chamber wall 125, the WSGP of FIG. 5 can be calibrated
by scaling the distance measured by the WSGP in the
direction of electrodes 60 and 62 to the calibrated
distances measured by MELP. Finally, since the
25 electrodes are contacting the chamber wall and providing

electrograms, the WEGP of FIG. 10 and SEGP of FIG. 14 can be calibrated to those measurements by computing the voltages at the same locations on the chamber wall 125 where electrodes 60 and 62 are located. These computed
5 voltages can then be scaled to match the physically measured voltages from electrodes 60 and 62.

THERAPY CATHETER

FIG. 16 is a schematic diagram of the therapy
10 catheter system. The therapy catheter 18 has both a distal end 24 and a proximal end 26. A handle 163 is on the proximal end 26 which allows the user to manipulate the distal end 24 and position it in the heart 16. Referring to FIG. 1, this handle also permits the
15 therapy catheter 18 to connect to the interface system 28 of the electrophysiologic apparatus 10 through the cable 32. The location current is generated by the signal generator 54 through the switch 59 and subsequently through the wire 177 of cable 32 which is
20 connected directly to the locator electrode 68. The therapy catheter system also includes a therapy generator 38 which is connected to the therapy catheter handle 163 via therapy supply line 161. The therapy supply line 161 extends through the handle 163, through
25 the catheter body 64, to the therapy deployment

apparatus 60 at the distal end 24 of the catheter. The locator electrode 68 is in close proximity to the therapy deployment apparatus 60 in order to determine its location within the heart 16.

5 FIG. 17 shows an embodiment of the therapy catheter 18 using laser energy to supply the therapy. This laser catheter 165 includes the location wire 177 which connects the interface system 28 to the locator electrode 68 at the catheter's distal end 24. In this
10 instance the therapy supply line 161 is a fiber optic cable 167 and the therapy deployment apparatus 60 is a fiber optic terminator 169 which directs the laser energy to the site of therapy delivery.

 FIG. 18 shows an embodiment of the therapy
15 catheter 18 using microwave energy to supply the therapy. This microwave catheter 171 includes the location wire 177 which connects the interface system 28 to the locator electrode 68 at the catheter's distal end 24. In this instance the therapy supply line 161 is a
20 microwave wave guide 173 and the therapy deployment apparatus 60 is a microwave emitter 175 which directs the microwave energy to the site of therapy delivery.

 FIG. 19 shows an embodiment of the therapy catheter 18 using a chemical to supply the therapy.
25 This chemical deliver catheter 181 includes the location

wire 177 which connects the interface system 28 to the locator electrode 68 at the catheter's distal end 24.

In this instance the therapy supply line 161 is a chemical filled lumen 183. This lumen extends to the

5 distal end 24 of the chemical delivery catheter 181
where a needle 185 is used to infuse the chemical into
the heart chamber wall 125. During introduction of the
chemical delivery catheter 181 into the heart chamber
the needle 185 is withdrawn into the catheter body

10 through withdrawal action 187. Once the location of the
distal end 24 is determined to be at the site of
interest the chemical delivery needle 185 can be
deployed through the reverse of withdrawal action 187.
Potential chemicals to be used in the therapeutic
15 delivery process include formaldehyde and alcohol.

Each of the therapy catheters 18 shown in FIG. 17 through FIG. 19 as well as the radio frequency catheter shown in FIG. 2 can be miniaturized and inserted into the coronary arterial tree. The location signal generated at locator electrode 68 can still be sensed by the passive electrodes 48 even though the signal is coming from the epicardium of the heart 16 rather than from within the heart chamber. Thus the movable electrode location process of FIG. 15 can be used in this instance to help determine the location of the

distal end 24 of the therapy catheter 18 in the coronary arterial tree and whether it is close to a site of abnormal electrical activity. Assuming that a site of ischemia will commonly be a site of abnormal electrical activity, the MELP will also enable more rapid location of potential sites for angioplasty.

FIG. 20 shows an embodiment of the therapy catheter 18 using balloon inflation to supply the therapy. This angioplasty catheter 191 includes the location wire 177 which connects the interface system 28 to the locator electrode 68 at the catheter's distal end 24. In this instance the therapy supply line 161 is an inflation media supply lumen 193 and the therapy deployment apparatus 60 is an angioplasty balloon 195. In use, a site of interest would be determined after viewing the wall electrogram generated by the WEGP of FIG. 10. Next the angioplasty therapy catheter 191 would be positioned in the coronary arterial tree and its position determined relative to the site of interest. Next, when the distal end 24 of the angioplasty catheter 191 was at the proper location the balloon 195 would be deployed to open the artery. Finally, the electrical activity of the site would be reviewed to determine whether the underlying tissue was now receiving a proper blood supply and thus was no

longer electrically abnormal.

002740" 00464560

We claim:

1. A therapy catheter comprising:
 - a) a lead body having a distal end and having a proximal end;
 - b) a locator electrode proximate said distal end;
 - c) a set of therapy electrodes located near said distal end.
2. A therapy catheter comprising:
 - a) a lead body having a distal end and having a proximal end;
 - b) a locator electrode proximate said distal end;
 - c) a drug delivery lumen located proximate said distal end.
3. A therapy catheter comprising:
 - a) a lead body having a distal end and having a proximal end;
 - b) a locator electrode proximate said distal end;
 - c) a fiber optic cable coupled to said distal end for directing laser energy to an ablation site.
4. A therapy catheter comprising:
 - a) a lead body having a distal end and having a proximal end;
 - b) a locator electrode proximate said distal end;
 - c) a microwave wave guide coupled to said distal end for directing microwave energy to an ablation site.
5. A therapy catheter comprising:
 - a) a lead body having a distal end and having a proximal end;
 - b) a locator electrode proximate said distal end;
 - c) an angioplasty balloon coupled to said distal end for opening a stenotic lesion in a coronary vessel.

ABSTRACT

A mapping catheter is positioned in a heart chamber, and active electrode sites are activated to impose an electric field within the chamber. The blood volume and wall motion modulates the electric field, which is detected by passive electrode sites on the preferred catheter. Electrophysiology measurements, as well as geometry measurements, are taken from the passive electrodes and used to display a map of intrinsic heart activity.




FIG. 1




FIG. 2




FIG. 3

FIG. 4




FIG. 5




FIG. 6




FIG. 7




FIG. 8

Place passive electrodes in a heart chamber & a pair of skin electrodes on patient

a) Frequency and timing of orientation current

a) Location of passive electrodes

(a) Wall surface generation process

a) Location of skin electrodes

Generate a current
between the pair
of skin electrodes

Measure voltage at passive electrodes resulting from current generation

Perform a regression to compute the vector of orientation current

Create a model of the heart chamber geometry

Generate a representation of chamber wall surface and electrode array

Generate a dynamic display of chamber wall movement in fluoroscopic like images

FIG. 9

Wall electrogram generation process

Knowledge provided to the process




FIG. 10




FIG. 11




FIG. 12




FIG. 13

Site electrogram generation process

Knowledge provided to the process




FIG. 14

Movable electrode location process

Knowledge provided to the process




FIG. 15




FIG. 16







FIG. 17



Please type a plus sign (+) inside this box → ☐

PTO/SB/01 (12-97)
Approved for use through 9/30/00 OMB 0651-0032
Patent and Trademark Office, U.S. DEPARTMENT OF COMMERCE

Under the Paperwork Reduction Act of 1995, no persons are required to respond to a collection of information unless it contains a valid OMB control number

DECLARATION FOR UTILITY OR DESIGN PATENT APPLICATION (37 CFR 1.63) <input type="checkbox"/> Declaration Submitted with Initial Filing OR <input type="checkbox"/> Declaration Submitted after Initial Filing (surcharge (37 CFR 1.16 (e)) required)	Attorney Docket Number	1930
	First Named Inventor	Beatty, G.
	COMPLETE IF KNOWN	
	Application Number	/
	Filing Date	4/12/00
	Group Art Unit	TBD
	Examiner Name	TBD

As a below named inventor, I hereby declare that:

My residence, post office address, and citizenship are as stated below next to my name.

I believe I am the original, first and sole inventor (if only one name is listed below) or an original, first and joint inventor (if plural names are listed below) of the subject matter which is claimed and for which a patent is sought on the invention entitled:

Electrophysiology Therapy Catheter

the specification of which (Title of the Invention)

☒ is attached hereto
OR
☐ was filed on (MM/DD/YYYY) [] as United States Application Number or PCT International Application Number [] and was amended on (MM/DD/YYYY) [] (if applicable)

I hereby state that I have reviewed and understand the contents of the above identified specification, including the claims, as amended by any amendment specifically referred to above

I acknowledge the duty to disclose information which is material to patentability as defined in 37 CFR 1.56.

I hereby claim foreign priority benefits under 35 U.S.C. 119(a)-(d) or 365(b) of any foreign application(s) for patent or inventor's certificate, or 365(a) of any PCT international application which designated at least one country other than the United States of America, listed below and have also identified below, by checking the box, any foreign application for patent or inventor's certificate, or of any PCT international application having a filing date before that of the application on which priority is claimed.

Prior Foreign Application Number(s)	Country	Foreign Filing Date (MM/DD/YYYY)	Priority Not Claimed	Certified Copy Attached?	
				YES	NO
			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>
			<input type="checkbox"/>	<input type="checkbox"/>	<input type="checkbox"/>

☐ Additional foreign application numbers are listed on a supplemental priority data sheet PTO/SB/02B attached hereto.

I hereby claim the benefit under 35 U.S.C. 119(e) of any United States provisional application(s) listed below.

Application Number(s)	Filing Date (MM/DD/YYYY)	<input type="checkbox"/> Additional provisional application numbers are listed on a supplemental priority data sheet PTO/SB/02B attached hereto.

[Page 1 of 2]

Burden Hour Statement: This form is estimated to take 0.4 hours to complete. Time will vary depending upon the needs of the individual case. Any comments on the amount of time you are required to complete this form should be sent to the Chief Information Officer, Patent and Trademark Office, Washington, DC 20231. DO NOT SEND FEES OR COMPLETED FORMS TO THIS ADDRESS. SEND TO: Assistant Commissioner for Patents, Washington, DC 20231

Please type a plus sign (+) inside this box → ☐

PTO/SB/01 (12-97)
Approved for use through 9/30/00 OMB 0651-0032
Patent and Trademark Office, U.S. DEPARTMENT OF COMMERCE

Under the Paperwork Reduction Act of 1995, no persons are required to respond to a collection of information unless it contains a valid OMB control number.

DECLARATION — Utility or Design Patent Application

I hereby claim the benefit under 35 U.S.C. 120 of any United States application(s), or 365(c) of any PCT international application designating the United States of America, listed below and, insofar as the subject matter of each of the claims of this application is not disclosed in the prior United States or PCT international application in the manner provided by the first paragraph of 35 U.S.C. 112, I acknowledge the duty to disclose information which is material to patentability as defined in 37 CFR 1.56 which became available between the filing date of the prior application and the national or PCT international filing date of this application

U.S. Parent Application or PCT Parent Number	Parent Filing Date (MM/DD/YYYY)	Parent Patent Number (if applicable)
09/005,105	1/9/1998	
08/387,832	2/16/1995	
PCT/US93/09015	9/23/1993	

☐ Additional U.S. or PCT international application numbers are listed on a supplemental priority data sheet PTO/SB/02B attached hereto.

As a named inventor, I hereby appoint the following registered practitioner(s) to prosecute this application and to transact all business in the Patent and Trademark Office connected therewith

☐ Customer Number

OR

☒ Registered practitioner(s) name/registration number listed below

Place Customer Number Bar Code Label here

Name	Registration Number	Name	Registration Number
Robert C. Beck	28,184		
Daniel A. Tysver	35,726		
Stephanie J. Smith	34,437		

☐ Additional registered practitioner(s) named on supplemental Registered Practitioner Information sheet PTO/SB/02C attached hereto

Direct all correspondence to: ☐ Customer Number or Bar Code Label ☐ OR ☒ Correspondence address below

Name	Beck & Tysver, P.L.L.P.		
Address	1011 First Street South, Suite 440		
Address			
City	Hopkins	State	MN
ZIP	55343		
Country	USA	Telephone	612-933-5042
Fax	612-933-3049		

I hereby declare that all statements made herein of my own knowledge are true and that all statements made on information and belief are believed to be true, and further that these statements were made with the knowledge that willful false statements and the like so made are punishable by fine or imprisonment, or both, under 18 U.S.C. 1001 and that such willful false statements may jeopardize the validity of the application or any patent issued thereon.

Name of Sole or First Inventor:		<input type="checkbox"/> A petition has been filed for this unsigned inventor		
Given Name (first and middle (if any))		Family Name or Surname		
Graydon		Beatty		
Inventor's Signature		Date		
Residence: City	St Paul	State	MN	
Country	USA		Citizenship	USA
Post Office Address	1935 Summit Ave			
Post Office Address				
City		State	MN	
ZIP	55105		Country	USA

☐ Additional inventors are being named on the _____ supplemental Additional Inventor(s) sheet(s) PTO/SB/02A attached hereto

Please type a plus sign (+) inside this box → ☐

PTO/SB/02A (3-97)
Approved for use through 9/30/98. OMB 0651-0032
Patent and Trademark Office, U.S. DEPARTMENT OF COMMERCE

Under the Paperwork Reduction Act of 1995, no persons are required to respond to a collection of information unless it contains a valid OMB control number

DECLARATION

ADDITIONAL INVENTOR(S)
Supplemental Sheet
Page ____ of ____

Name of Additional Joint Inventor, if any:				<input type="checkbox"/> A petition has been filed for this unsigned inventor				
Given Name (first and middle [if any])				Family Name or Surname				
Jeffrey				Budd				
Inventor's Signature				Date				
Residence: City		St Paul	State	MN	Country	USA	Citizenship	USA
Post Office Address		2261 Gordon Ave.						
Post Office Address								
City		St Paul	State	MN	ZIP	55112	Country	USA
Name of Additional Joint Inventor, if any:				<input type="checkbox"/> A petition has been filed for this unsigned inventor				
Given Name (first and middle [if any])				Family Name or Surname				
John				Hauck				
Inventor's Signature				Date				
Residence: City		Shoreview	State	MN	Country	USA	Citizenship	USA
Post Office Address		5900 Hodgson Road						
Post Office Address								
City		Shoreview	State	MN	ZIP	55126	Country	USA
Name of Additional Joint Inventor, if any:				<input type="checkbox"/> A petition has been filed for this unsigned inventor				
Given Name (first and middle [if any])				Family Name or Surname				
Inventor's Signature				Date				
Residence: City			State		Country		Citizenship	
Post Office Address								
Post Office Address								
City			State		ZIP		Country	

Burden Hour Statement: This form is estimated to take 0.4 hours to complete. Time will vary depending upon the needs of the individual case. Any comments on the amount of time you are required to complete this form should be sent to the Chief Information Officer, Patent and Trademark Office, Washington, DC 20231. DO NOT SEND FEES OR COMPLETED FORMS TO THIS ADDRESS. SEND TO Assistant Commissioner for Patents, Washington, DC 20231

PTO/SB/02B (3-97)
Approved for use through 9/30/98. OMB 0651-0032
Patent and Trademark Office, U.S. DEPARTMENT OF COMMERCE
re required to respond to a collection of information unless it contains a

Additional foreign applications:

[illegible]

Additional provisional applications:

Application Number	Filing Date (MM/DD/YYYY)

Additional U.S. applications:

U.S. Parent Application Number	PCT Parent Number	Parent Filing Date (MM/DD/YYYY)	Parent Patent Number (if applicable)
07/950,448		9/23/1992	5,291,549
07/949,690		9/23/1992	5,311,866

Burden Hour Statement This form is estimated to take 0.4 hours to complete. Time will vary depending upon the needs of the individual case. Any comments on the amount of time you are required to complete this form should be sent to the Chief Information Officer, Patent and Trademark Office, Washington, DC 20231. DO NOT SEND FEES OR COMPLETED FORMS TO THIS ADDRESS. SEND TO: Assistant Commissioner for Patents, Washington, DC 20231

Parameter	Value	Unit
Temperature (°C)	25.0	°C
Pressure (kPa)	101.3	kPa
Humidity (%)	65.0	%
Wind speed (m/s)	1.5	m/s
Wind direction (°)	180	°
Cloud cover (%)	10	%
Relative humidity (%)	65.0	%
Atmospheric pressure (hPa)	1013.25	hPa
Sea level pressure (hPa)	1013.25	hPa
Barometric pressure (hPa)	1013.25	hPa
Altitude (m)	0	m
Latitude (°N)	40.71	°N
Longitude (°W)	84.04	°W
Time (h:m:s)	12:00:00	h:m:s
Date (dd/mm/yyyy)	15/08/2018	dd/mm/yyyy
Location (City/Country)	New York/USA	City/Country
Observer (Name)	John Doe	Name
Instrument (Model)	Model X-100	Model
Calibration (Date)	01/07/2018	Date
Notes	Clear sky, light breeze.	Text